



CERTIFICATE OF ANALYSIS

Customer : Garden State Environmental
555 South Broad Street, Suite K
Glen Rock, NJ 07452

Project ID : Oradell #6409
PAS Project ID : P17-1639

Matrix : Drinking Water
Report Date : 4/26/2017

PAS Sample ID	Client ID	Analysis	Results	Units	DF	PQL	MDL	MCL	Method	Date Sampled	Date Analyzed
P17-1639-01	#1A	Lead	2.41	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 05:57	4/19/17 10:31
P17-1639-02	#2A	Lead	1.17 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:00	4/19/17 10:38
P17-1639-03	#3A	Lead	1.41 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:02	4/19/17 10:49
P17-1639-04	#4A	Lead	0.667 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:03	4/19/17 11:01
P17-1639-05	#5A	Lead	1.17 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:07	4/19/17 11:04
P17-1639-06	#29A	Lead	2.16	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:12	4/19/17 11:08
P17-1639-07	#30A	Lead	1.91 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:18	4/19/17 11:12
P17-1639-08	#6A	Lead	1.17 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:21	4/19/17 11:15
P17-1639-09	#7A	Lead	1.17 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:22	4/19/17 11:19
P17-1639-10	#8A	Lead	9.39	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:23	4/19/17 11:23
P17-1639-11	#9A	Lead	4.90	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:30	4/19/17 11:27
P17-1639-12	#10A	Lead	4.41	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:35	4/19/17 11:31
P17-1639-13	#11A	Lead	1.17 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:39	4/19/17 11:42
P17-1639-14	#12A	Lead	1.41 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:41	4/19/17 11:46
P17-1639-15	#13A	Lead	1.41 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:44	4/19/17 11:50
P17-1639-16	#14A	Lead	1.17 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:46	4/19/17 11:54
P17-1639-17	#15A	Lead	0.667 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:48	4/19/17 11:58
P17-1639-18	#16A	Lead	0.667 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:57	4/19/17 12:02
P17-1639-19	#17A	Lead	0.916 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 06:59	4/19/17 12:05
P17-1639-20	#18A	Lead	1.41 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 07:04	4/19/17 12:09
P17-1639-21	#19A	Lead	4.16	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 07:08	4/19/17 12:13
P17-1639-22	#31A	Lead	9.39	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 07:09	4/19/17 12:28
P17-1639-23	#20A	Lead	2.16	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 07:11	4/19/17 12:43
P17-1639-24	#21A	Lead	3.91	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 07:14	4/19/17 12:47
P17-1639-25	#23A	Lead	1.17 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 07:18	4/19/17 12:51
P17-1639-26	#24A	Lead	1.41 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 07:22	4/19/17 12:54
P17-1639-27	#25A	Lead	0.916 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 07:23	4/19/17 12:58
P17-1639-28	#26A	Lead	1.41 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 07:26	4/19/17 13:09
P17-1639-29	#22A	Lead	16.8	ug/L	2	4.00	0.850	15.0 *	SM 3113 B	4/10/17 07:32	4/19/17 14:50
P17-1639-30	#27A	Lead	0.667 J	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 07:39	4/19/17 13:17
P17-1639-31	#28A	Lead	ND	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 07:40	4/19/17 13:20
P17-1639-32	#32A	Lead	ND	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 07:47	4/19/17 13:24
P17-1639-33	ORA-4-10-FBA	Lead	ND	ug/L	1	2.00	0.425	15.0 *	SM 3113 B	4/10/17 07:50	4/19/17 13:28

Except for the parameters tested, PAS makes no representation as to the fitness or quality of the water sample taken.

- PQL = Practical Quantitation Limit
- MDL = Minimum Detection Limit
- MCL = Maximum Contaminant Level
- DF = Dilution Factor
- ND = Analyzed for but not detected
- J = Estimated result
- * Federal Action Level

All samples are analyzed in accordance with New Jersey Department of Environmental Protection Protocol

Mark D. Feitelson, Lab. Director